

CALIFORNIA LABORATORY SERVICES

3249 Fitzgerald Road Rancho Cordova, CA 95742

April 15, 2016

CLS Work Order #: CZD0654
COC #: 169788


Rob Kull
Stratus Environmental
3330 Cameron Park Dr. #550
Cameron Park, CA 95682

Project Name: Mangan Park Gun Range

Enclosed are the results of analyses for samples received by the laboratory on 04/14/16 09:12. Samples were analyzed pursuant to client request utilizing EPA or other ELAP approved methodologies. I certify that the results are in compliance both technically and for completeness.

Analytical results are attached to this letter. Please call if we can provide additional assistance.

Sincerely,



James Liang, Ph.D.
Laboratory Director

CA DOHS ELAP Accreditation/Registration number 1233

CALIFORNIA LABORATORY SERVICES

Page 2 of 4

04/15/16 14:09

Stratus Environmental
3330 Cameron Park Dr. #550
Cameron Park, CA 95682

Project: Mangan Park Gun Range
Project Number: [none]
Project Manager: Rob Kull

CLS Work Order #: CZD0654
COC #: 169788

Metals by EPA 6000/7000 Series Methods

Analyte	Result	Reporting Limit	Units	Dilution	Batch	Prepared	Analyzed	Method	Notes
Pool Water (CZD0654-01) Water Sampled: 04/14/16 08:09 Received: 04/14/16 09:12									
Lead	ND	5.0	µg/L	1	CZ02661	04/15/16	04/15/16	EPA 6020	

CALIFORNIA LABORATORY SERVICES

Stratus Environmental 3330 Cameron Park Dr. #550 Cameron Park, CA 95682	Project: Mangan Park Gun Range Project Number: [none] Project Manager: Rob Kull	CLS Work Order #: CZD0654 COC #: 169788
---	---	--

Metals by EPA 6000/7000 Series Methods - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch CZ02661 - EPA 3020A

Blank (CZ02661-BLK1)				Prepared & Analyzed: 04/15/16						
Lead	ND	5.0	µg/L							
LCS (CZ02661-BS1)				Prepared & Analyzed: 04/15/16						
Lead	103	5.0	µg/L	100		103	80-120			
Matrix Spike Dup (CZ02661-MSD1)				Source: CZD0608-03 Prepared & Analyzed: 04/15/16						
Lead	104	5.0	µg/L	100	0.260	104	75-125	25		QR-2

CALIFORNIA LABORATORY SERVICES

Page 4 of 4

04/15/16 14:09

Stratus Environmental
3330 Cameron Park Dr. #550
Cameron Park, CA 95682

Project: Mangan Park Gun Range
Project Number: [none]
Project Manager: Rob Kull

CLS Work Order #: CZD0654
COC #: 169788

Notes and Definitions

- QR-2 The RPD result exceeded the QC control limits; however, both percent recoveries were acceptable. Sample results for the QC batch were accepted based on percent recoveries and completeness of QC data.
- DET Analyte DETECTED
- ND Analyte NOT DETECTED at or above the reporting limit (or method detection limit when specified)
- NR Not Reported
- dry Sample results reported on a dry weight basis
- RPD Relative Percent Difference